

DoD Handbook for Perchlorate Sampling and Testing

EPA Region 6

14th Annual Quality Assurance Conference
October 2004


Edward Hartzog

DoD Environmental Data Quality Workgroup
Navy Laboratory Quality & Accreditation Office

tabbles®

EXHIBIT

G



DOD Policy

Interim Guidance on Sampling and Testing for Perchlorate (Feb 2004)

- Laboratories shall comply with the DoD QSM
- Results above the regulatory limit must be verified using definitive technology
- Methods other than 314.0 must be used if target MRL < 4ppb
- Ability to achieve MRL must be demonstrated in matrix of concern
- QA/QC for alternative methods must be equivalent to or more stringent than those of Method 314.0