

Reported Device Return Rate and Reason for Out of Service

In accordance with HRS-draft recommended elements in manufacturer product performance reports, Boston Scientific provides the reason(s) for U.S. device explant or out of service if known, for all pulse generator products/product groupings for which U.S. survival probability data are reported. Table 1 also quantifies the number of Boston Scientific devices reported to have been explanted in the U.S., and the number and percentage of explanted devices then returned to Boston Scientific for various product types. The reasons for device explant or out of service consist of normal battery depletion, device upgrade, device malfunction (which includes devices under advisory that have experienced a malfunction), complication related to another system component or clinical conditions such as infection, or "other," a category consisting of patient death, prophylactic device explant, elective replacement, general product dissatisfaction, other observation/complication, unspecified, or unknown. It should be noted that the reason for a device explant or out of service may either be confirmed through laboratory analysis (as in the case of normal battery depletion) or it may be reported to Boston Scientific with no associated returned device or laboratory analysis. Additionally, there may be more than one reason listed for why a device was explanted/taken out of service.

Reported Device Return Rate

U. S. Implant Population (All-Time)

Product Type	Explanted devices reported to Boston Scientific CRM	Reported device return rate
CRTDs and CRT-Ps	8,436	5,678 (66%)
ICDs	75,634	48,765 (66%)
Pacemakers	59,235	37,719 (64%)
Total	143,305	93,062 (65%)

Table 1. Explanted device return rate to Boston Scientific CRM.

Reason for Out of Service

CRT-Ds/Model	Total U.S. Registered Implants	Normal Battery Depletion	Device Upgrade	Device Malfunction ¹	Complication related to another system component or a clinical condition ²	Other ³
CRT-Ds/Model						
CONTAK RENEWAL 3 RF HE H217/H219	3,000	0	0	0	17	94
CONTAK RENEWAL 3 RF H210/H215	4,000	0	0	1	21	93
CONTAK RENEWAL 3 HE H177/H179	21,000	38	1	27	251	3,532
CONTAK RENEWAL 3 H170/H176	30,000	41	1	46	298	4,827
CONTAK RENEWAL H135	10,000	665	2	92	156	5,004
CRT-Ps/Model						
CONTAK RENEWAL TR H120/H125	5,000	16	13	0	34	1,020

¹Device malfunction consists of all U.S. confirmed malfunctions for a product/product grouping. These include confirmed malfunctions for advisory populations, as well as any other type of malfunction in which a device was returned and confirmed by laboratory analysis to have malfunctioned. U.S. confirmed malfunction counts are reflected in the U.S. Summary tables and in U.S. survival probability data.
²System component and/or clinical condition complications may include, for example, infection, erosion, lead-to-PG interface.
³Other consists of: patient death, elective replacement, general product dissatisfaction, other observation/complication, unspecified, or unknown.
⁴Counts consist of Boston Scientific pacemaker and Intermedica co-branded pacemaker data.

continued on next page

REASON FOR OUT OF SERVICE

ICDs/Model	Total U.S. Registered Implants	Normal Battery Depletion	Device Upgrade	Devices Malfunction ¹	Complication related to another system component or a clinical condition ²	Other ³
VITALITY 2 EL VR T177	3,000	0	3	3	13	132
VITALITY 2 EL DR T187	3,000	1	4	2	20	151
VITALITY 2 VR T176	9,000	1	11	19	49	609
VITALITY 2 DR T185	13,000	4	9	18	64	770
VITALITY DR HE T180	3,000	2	1	0	15	111
VITALITY AVT A155	12,000	10	17	47	100	1,914
VITALITY AVT A135	7,000	37	21	69	70	1,780
VITALITY DS DR T125	16,000	7	18	24	105	1,877
VITALITY DS VR T135	14,000	9	26	19	82	1,785
VITALITY EL T127	3,000	0	2	4	26	232
VENTAK PRIZM 2 DR 1861	43,000	843	148	160	495	19,286
VENTAK PRIZM 2 VR 1860	25,000	108	101	57	226	7,423
VENTAK PRIZM DR HE 1853/1858	10,000	1,711	23	78	136	3,040
VENTAK PRIZM VR HE 1852/1857	8,000	678	27	33	108	2,246
VENTAK PRIZM DR 1851/1856	13,000	4,479	44	260	166	5,698
VENTAK PRIZM VR 1850/1855	8,000	2,235	40	77	90	3,349

¹Device malfunction consists of all U.S. confirmed malfunctions for a product/product grouping. These include confirmed malfunctions for advisory populations, as well as any other type of malfunction in which a device was returned and confirmed by laboratory analysis to have malfunctioned. U.S. confirmed malfunction counts are reflected in the U.S. Summary tables and in U.S. survival probability data.
²System component and/or clinical condition complications may include, for example: infection, erosion, lead-to-PG interface.
³Other consists of: patient death, elective replacement, general product dissatisfaction, other observation/complication, unspecified, or unknown.
⁴Counts consist of Boston Scientific pacemaker and Intermedica co-branded pacemaker data.

Total U.S. Registered Implants

Pacemakers/Model

Pacemakers/Model	Total U.S. Registered Implants	Normal Battery Depletion	Device Upgrade	Device Malfunction ¹	Complication related to another system component or a clinical condition ²	Other ³
INSIGNIA Ultra SR ¹	10,000	1	7	0	28	1,204
INSIGNIA Ultra DR downsize ¹	33,000	3	18	3	99	2,499
INSIGNIA Ultra DR ¹	12,000	1	8	1	39	625
INSIGNIA Entra SR ¹	9,000	9	6	0	28	2,280
INSIGNIA Entra DR downsize ¹	18,000	14	14	4	63	3,144
INSIGNIA Entra DR ¹	12,000	4	2	6	44	1,890
INSIGNIA Plus SR ¹	23,000	32	38	6	77	6,942
INSIGNIA Plus DR ¹	21,000	39	31	3	105	3,265
INSIGNIA Plus DR downsize ¹	77,000	279	88	33	368	15,422
PULSAR MAX II SR downsize ¹	7,000	49	14	1	25	3,149
PULSAR MAX II SR ¹	1,000	4	4	0	5	511
PULSAR MAX II DR ¹	29,000	261	86	13	152	8,489
DISCOVERY II SR downsize ¹	13,000	110	15	2	63	6,408
DISCOVERY II SR ¹	3,000	17	7	1	15	1,488
DISCOVERY II DR downsize ¹	33,000	2,638	55	17	178	13,336
DISCOVERY II DR ¹	23,000	264	30	8	120	7,542
PULSAR MAX SR downsize ¹	11,000	580	28	6	46	6,188
PULSAR MAX SR ¹	2,000	30	16	6	11	1,344
PULSAR MAX DR ¹	41,000	2,660	110	79	246	17,893

¹Device malfunction consists of all U.S. confirmed malfunctions for a product/product grouping. These include confirmed malfunctions for advisory populations, as well as any other type of malfunction in which a device was returned and confirmed by laboratory analysis to have malfunctioned. U.S. confirmed malfunction counts are reflected in the U.S. Summary tables and in U.S. survival probability data.

²System component and/or clinical condition complications may include, for example: infection, erosion, lead-to-PG interface.

³Other consists of: patient death, elective replacement, general product dissatisfaction, other product dissatisfaction, other observation/complication, unspecified, or unknown.

⁴Counts consist of Boston Scientific pacemaker and Intermedics co-branded pacemaker data.

continued on next page

REASON FOR OUT OF SERVICE

	Total U.S. Registered Implants	Normal Battery Depletion	Device Upgrade	Device Malfunction ¹	Complication related to another system component or a clinical condition ²	Other ³
DISCOVERY SR downsize ⁴ 1174	10,000	531	28	78	48	6,188
DISCOVERY SR ⁵ 1175	3,000	43	6	1	20	1,619
DISCOVERY DR downsize ⁴ 1273	23,000	6,065	40	147	119	13,167
DISCOVERY DR ⁵ 1274/1275	13,000	1,103	30	119	91	7,325
MERIDIAN DDD 0976	2,000	35	0	4	13	1,040
MERIDIAN SR 1176	9,000	200	15	18	38	5,584
MERIDIAN DR 1276	16,000	527	18	35	98	8,025

¹Device malfunction consists of all U.S. confirmed malfunctions for a product/product grouping. These include confirmed malfunctions for advisory populations, as well as any other type of malfunction in which a device was returned and confirmed by laboratory analysis to have malfunctioned. U.S. confirmed malfunction counts are reflected in the U.S. Summary tables and in U.S. survival probability data.

²System component and/or clinical condition complications may include, for example: infection, erosion, lead-to-PG interface.

³Other consists of: patient death, elective replacement, general product desaturation, other observation/complication, unspecified, or unknown.

⁴Counts consist of Boston Scientific pacemaker and Intermedics co-branded pacemaker data.

U.S. Reports of Acute Lead Observations (Occurring Within First Month of Service)

Boston Scientific strives to provide meaningful product performance information reflective of real-world product experience. To provide more meaningful lead cumulative survival information, Boston Scientific's lead statistics include all reports of lead malfunctions, whether or not confirmed by laboratory analysis. A lead complication is said to have occurred if any of the following conditions have been met:

- Laboratory analysis of a returned lead determines that the lead was not compliant in form, fit or function at any time while implanted
- Any clinical observation related to a discrepancy with the mechanical integrity and/or impedance measurement of the lead, 30 days or more post-implant.

Lead complications as defined above are included in the lead cumulative survival graphs in previous sections. Because acute lead performance contributes to overall clinical experience, Boston Scientific provides a summary of U.S. reports of lead observations occurring in the first month of service. Acute lead performance may be subject to a number of factors, including patient-specific clinical conditions and/or varying implant conditions/techniques. To be included in this summary of observations, a lead must first be successfully implanted. These reports may or may not have resulted in clinical action and/or product return to Boston Scientific. Multiple observations are possible for any given lead.

Total U.S. Registered Implants

Pacing Leads/Model	Total U.S. Registered Implants	Impedance ¹	Physical Damage/ Mechanical Connection ²	Dislodgement	Cardiac Perforation	Pacing Capture/ Threshold Issue	Muscle Stimulation	Sensing/ Detection ³	Noisy Signal Observed	High DFT	Therapy Delivery/ Effectiveness ⁴
FLEXTEND	137,000	177	268	695	105	888	48	259	19	0	47
Active Fixation 4086/4087/4088											
FINELINE II/FINELINE II Sterox¹	90,000	25	39	137	4	162	6	39	8	0	8
Passive Fixation (polyurethane) 4462/4463/4464/4465/4466/4467											
FINELINE II EZ/FINELINE II Sterox EZ¹	174,000	116	121	213	16	227	22	104	11	1	30
Positive Fixation (polyurethane) 4463/4464/4465/4466/4467/4471											
FINELINE II/FINELINE II Sterox¹	28,000	13	33	196	0	97	2	58	5	0	7
Atrial J (polyurethane) 4477/4478/4479/4480											
FINELINE II/FINELINE II Sterox¹	10,000	5	12	14	1	16	1	8	2	0	1
Passive Fixation (silicone) 4454/4455/4456/4459											
FINELINE II EZ/FINELINE II Sterox EZ¹	29,000	9	39	54	2	34	5	19	4	0	1
Positive Fixation (silicone) 4466/4467/4468/4472/4473/4474											

CRT Leads/Model	Total U.S. Registered Implants	Impedance ¹	Physical Damage/ Mechanical Connection ²	Dislodgement	Vein Dissection	Pacing Capture/ Threshold Issue	Muscle Stimulation	Sensing/ Detection ³	Noisy Signal Observed	High DFT	Therapy Delivery/ Effectiveness ⁴
EASYTRAK 3	7,000	8	3	68	2	40	49	6	1	0	1
4549/4549/4550/4522/4524/4525/4527											
EASYTRAK 2	29,000	83	24	245	11	154	245	25	2	0	18
4542/4543/4544/4515/4517/4518/4520											
EASYTRAK	35,000	43	39	205	6	205	203	28	6	0	17
4537/4538/4510/4511/4512/4513											

continued on next page

U.S. REPORTS OF ACUTE LEAD OBSERVATIONS

Defibrillation Leads/Model	Total U.S. Registered Implants	Impedance ¹	Physical Damage/ Mechanical Connection?	Dislodgement	Cardiac Perforation	Pacing Capture/ Threshold Issue	Muscle Stimulation	Sensing/ Detection ²	Noisy Signal Observed	High DFT	Therapy Delivery/ Effectiveness ⁴
ENDOTAK RELIANCE G	52,000	59	82	114	19	110	12	83	54	20	61
Dual Coil, Active Fixation 0184/0185/0186/0187											
ENDOTAK RELIANCE G	5,000	19	0	11	1	21	0	6	4	1	4
Dual Coil, Passive Fixation 0174/0175/0179/0177											
ENDOTAK RELIANCE SG	2,000	4	5	1	2	3	1	1	1	2	1
Single Coil, Active Fixation 0180/0181/0182											
ENDOTAK RELIANCE	70,000	78	94	81	18	102	11	121	77	22	67
Dual Coil, Active Fixation 0157/0158/0159											
ENDOTAK RELIANCE	30,000	60	6	55	3	79	2	48	30	7	34
Dual Coil, Passive Fixation 0147/0148/0149											
ENDOTAK RELIANCE S	1,000	3	2	0	0	0	0	2	1	1	1
Single Coil, Active Fixation 0137/0138											
ENDOTAK RELIANCE S	1,000	0	0	1	0	1	1	2	0	0	0
Single Coil, Passive Fixation 0127/0128											
ENDOTAK ENDURANCE EZ	29,000	21	36	53	2	44	5	177	11	8	54
Active Fixation 0154/0155/0156											
ENDOTAK ENDURANCE RX	18,000	21	1	32	0	21	1	30	2	2	18
Passive Fixation Steroid Eluting 0144/0145/0146											
ENDOTAK ENDURANCE	3,000	2	0	2	0	9	0	5	0	1	3
Passive Fixation 0134/0135/0136											
ENDOTAK DSP	36,000	11	18	36	0	26	4	79	1	21	30
Passive Fixation 0094/0095/0125											

¹Impedance: Low impedance, high impedance, unspecified impedance issue, or shorted lead.
²Physical Damage/Mechanical Connection: Damage to electrode, terminal, conductor, insulation, or lead insertion difficulty.
³Sensing/Detection: Oversensing, undersensing, intermittent sensing, or non-detection.
⁴Therapy Delivery/Effectiveness: Tachy or brady therapy either not delivered when required, or delivered when not required; non-conversion; or configuration change required.
⁵Counts consist of Boston Scientific pacing lead and Intermedics pacing lead co-branded data.

Product Index

CONTAK RENEWAL 3 RF HE.....	7, 41
CONTAK RENEWAL 3 RF.....	7, 41
CONTAK RENEWAL 3 HE.....	8, 41
CONTAK RENEWAL.....	9, 42
CONTAK RENEWAL 3.....	8, 42
CONTAK RENEWAL TR.....	8, 40
DISCOVERY DR.....	36, 56
DISCOVERY DR (downsize).....	35, 56
DISCOVERY SR.....	35, 55
DISCOVERY SR (downsize).....	34, 55
DISCOVERY II DR.....	32, 56
DISCOVERY II DR (downsize).....	32, 53
DISCOVERY II SR.....	31, 53
DISCOVERY II SR (downsize).....	31, 53
EASYTRAK.....	11, 61
EASYTRAK 2.....	11, 61
EASYTRAK 3.....	11, 61
ENDOTAK ENDURANCE EZ.....	22, 64
ENDOTAK ENDURANCE Rx.....	22, 64
ENDOTAK ENDURANCE.....	23, 65
ENDOTAK 60 Series.....	24, 66
ENDOTAK 70 Series.....	24, 66
ENDOTAK DSP.....	23, 65
ENDOTAK PLUS.....	23, 65
ENDOTAK RELIANCE G Dual Coil.....	20, 62
ENDOTAK RELIANCE SG Single Coil.....	20, 62
ENDOTAK RELIANCE Dual Coil.....	21, 63
ENDOTAK RELIANCE S Single Coil.....	23, 64, 63

ENDURANCE products (see ENDOTAK ENDURANCE)	
FINELINE II EZ Polyurethane.....	39, 68
FINELINE II Sterox Polyurethane.....	38, 67
FINELINE II Sterox EZ Polyurethane.....	38, 67
FINELINE II Atrial J Polyurethane.....	39
FINELINE II Sterox Atrial J Polyurethane.....	37, 66
FINELINE II Sterox Silicone.....	39, 68
FINELINE II Sterox EZ Silicone.....	39, 68
FLEXTEND Active Fixation.....	38, 67
INSIGNIA Entra DR.....	27, 50
INSIGNIA Entra DR (downsize).....	27, 50
INSIGNIA Entra SR.....	26, 50
INSIGNIA Plus DR.....	28, 51
INSIGNIA Plus DR (downsize).....	29, 51
INSIGNIA Plus SR.....	28, 51
INSIGNIA Ultra DR.....	26, 49
INSIGNIA Ultra DR (downsize).....	25, 49
INSIGNIA Ultra SR.....	25, 49
MERIDIAN DDD.....	36, 56
MERIDIAN DR.....	37, 57
MERIDIAN SR.....	37, 57
PRIZM products (see VENTAK PRIZM)	
PULSAR MAX DR.....	34, 55
PULSAR MAX SR.....	33, 54
PULSAR MAX SR (downsize).....	33, 54

PULSAR MAX II DR.....	30, 52
PULSAR MAX II SR.....	30, 52
PULSAR MAX II SR (downsize).....	29, 52
RELIANCE products (see ENDOTAK RELIANCE)	
RENEWAL products (see CONTAK RENEWAL)	
VENTAK PRIZM DR.....	19, 48
VENTAK PRIZM DR HE.....	18, 47
VENTAK PRIZM VR.....	19, 48
VENTAK PRIZM VR HE.....	18, 47
VENTAK PRIZM 2 DR.....	17, 46
VENTAK PRIZM 2 VR.....	17, 46
VITALITY 2 DR.....	13, 43
VITALITY 2 EL DR.....	12, 43
VITALITY 2 EL VR.....	12, 44
VITALITY 2 VR.....	13, 44
VITALITY AVT.....	17, 44, 45
VITALITY DR HE.....	14, 43
VITALITY DS DR.....	15, 45
VITALITY DS VR.....	16, 45
VITALITY EL.....	16, 46

Trademarks.....	6
Confirmed Malfunction Details: Pulse Generators.....	40
Overview.....	411
CRT-D and CRT-P.....	43
ICDs.....	49
Pacemakers.....	58
References.....	61
Confirmed Malfunction Details: Leads CRT Leads.....	62
Defibrillation Leads.....	67
Pacing Leads.....	69
References.....	70
Safety Advisories.....	85
Reported device return rate and reason for out of service.....	89
U.S. Reports of Acute Lead Observations.....	92
Contact Information.....	

Contact Information

Guidant Europe S.A.

Park Lane
Culliganen 2B
1831 Diegem
Belgium
Tel: 32.2.714.14.11
Fax: 32.2.714.14.12

Argentina

Guidant Argentina
Cardiac Pacemakers, Inc.
Av. Pte. Roque Sáenz Peña 832 1st floor
1388 Buenos Aires
Tel: 54.11.4320.4500

Australia

Guidant Australia Pty Ltd.
Level 1, 4 Inglewood Place
Baulkham Hills NSW 2153
Australia
Tel: +61.2.9421.2800
Fax: +61.2.9421.2805

Austria

Guidant Austria GmbH
Vienna Twin Tower
Turm A / 19.0G
Wienerbergstraße 11
1100 Wien
Tel: +43.1.641.8800
Fax: +43.1.641.8800-10

Belgium

Guidant Belgium
Park Lane
Culliganen 2B
1831 Diegem
Tel: 32.2.714.16.11

Canada

Guidant Canada
505 Apple Creek Blvd. Unit 4
Suite 2201
Markham, Ontario L3R 5B1
Canada
Tel: 905.947.5600
Fax: 905.947.5630

Czech Republic

Guidant International B.V. Czech
Zeleny Pruh 99
14002 Praha 4
Tel: 420.2.96374700

Denmark

Guidant Denmark A/S
Lygten 2 C, 4
DK-2400 Copenhagen NV
Tel: 45.35.8700.10

France

Guidant France SA
9 Rue d'Estienne d'Orves
F-92504 Rueil Malmaison
Tel: 33.1.47.14.4014
www.guidant.fr

Germany

Guidant GmbH
Wingershecke 6
D-35392 Giessen
Tel: +49.641.92221.0
Fax: +49.641.92221.608
www.guidant.de

Hong Kong

Guidant Hong Kong
Suite 2201
CRC Mess Mutual Tower
3B Gloucester Road
Wan Chai, Hong Kong
Tel: 852.28272338

Italy

Guidant Italia, S.r.l.
Via Cassanese, 224
Palazzo Raffaello
I-20090 Segrate (MI)
Tel: 39.02.269831
www.guidant.it

Japan

Guidant Japan KK
Shinagawa East One Tower 10F
2-16-1, Konan, Minato-ku
Tokyo 108-0075
Japan
Tel: +81.3.67170100
Fax: +81.3.67170151/0251

The Netherlands

Guidant Nederland B.V.
Kulfrass 56
NL-3435 RG Nieuwegein
Tel: 31.30.802.55.55

New Zealand

Guidant New Zealand
103 Carleton Gore Road, Ground Floor
Newmarket
Auckland
Tel: 64.9.622.9900

Singapore

Guidant Singapore Pte Ltd.
152 Beach Road
#21-07/08 Gateway East
Singapore 189721
Tel: 65.392.0032

Portugal

Guidant Aparelhos Medicos Lda
Av' do Forte, nº 3
Edifício Suécia III, Piso -1
2795 - 504 Camaxide
Portugal
Tel: +351.21.416.4720

Spain

Guidant, SA
Ctra. Fluencarral
Alcobendas, Km. 15,700
Edif. Europa, nº 1, 3ª planta
E-28108 Alcobendas, Madrid
Tel: 34.91.65725.45

Sweden

Guidant Sweden AB
Sjölingsvägen 11
S-102 72 Sollentuna
Tel: 46.0.8.594.785.50

Switzerland

Guidant AG
Bundesplatz 1
CH-6300 Zug
Tel: +41.41.728.7788
Fax: +41.41.287780

United Kingdom

Guidant Ltd.
Hampshire International
Business Park
Crockford Lane
Chineham, Basingstoke
Hampshire UK-RG24 8WH
Tel: 44.1256.374.000

United States

Boston Scientific Corporation
4100 Hermine Avenue North
St. Paul, MN 55112-5798 USA
Tel: 651.582.4000
Fax: 651.582.4168
Medical Professionals:
1.800.CARDIAC (227.3422)
Patients and Families:
1.866.484.3268

For more information, visit our website at www.guidant.com